

Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

> TEL: +82-31-300-0400 / FAX: +82-31-300-0499 http://www.sdbiosensor.com

Certificate of Analysis

Product Name	STANDARD™ F NT-proBNP FIA			
Cat No.	10NTP10B	Ref No.	F-NTP-01	
LOT No.	6414672AC	MFG Date	2023.06.07.	
		EXP Date	2024.12.06.	
Testing categories		Specification		Result
Appearance Inspection	 Uniform Gap between upper and lower device No scratch, stain and crack etc. No contamination There should be no colored lines. Check device type Check buffer amount 			Pass
Barcode Check	The product name and 'apply sample' should be displayed on the screen.			Pass
Repeatability	• Whole blood, Serum CV ≤ 15%			Pass
Dynamic range	 Testing temperature range: 15-32°C 59-90°F Testing Range: 50 ~ 25,000 pg/mL 			Pass
Packaging Inspection	20 Test Device, 20 Buffer, Disposable dropper 100µl(20ea), 1 IFU			Pass
	No leackage in sealing part of pouch			Pass
	No leackage in sealing part of buffer solution tube			Pass
	Manufacturing and expiration date between the pouch and package of test kit.			Pass
	LOT No. between the pouch an	LOT No. between the pouch and package of test kit.		

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.



Ko Jongkwan

Quality Management Representative QA Division

Ko Jongkwan Quality Management Representative QA Division