F-QP-0026-0016-0077



Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA Manufacturing site: 14, Jeungpyeongsandan-ro, Jeungpyeong-eup, Jeungpyeong-gun, Chungcheongbuk-do, 27915, REPUBLIC OF KOREA

> TEL : +82-31-300-0400 / FAX : +82-31-300-0499 http://www.sdbiosensor.com

> > Rev.00 (2023.10.16.)

Certificate of Analysis

Product Name	STANDARD™ M10 SARS-CoV-2			
Cat No.	11COV10A	Ref No.	M-NCOV-03	
Lot No.	8C84CM1AJ1	MFG Date	2023-12-21	
		EXP Date	2024-12-20	
Testing categories	Specification		Result	
Performance Inspection	 Sensitivity: 100% detected in 2X LoD (1.33x10⁻³ TCID₅₀/ml ≒ 200 copies/ml) 			Pass
	Repeatability: Results are within ± 2 Ct of target concentration			Pass
	Specificity: No fluorescence signal detected in negative control			Pass
Apperance Inspection	 Check the text number of the labeling material Quantity per one container : 10ea 			Pass
	 Expiration date between cartridge label and package of test kit LOT No. between cartridge label and package of test kit 			Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.



Ko Jongkwan Quality Management Representative QA Division

Jongkwan Ko Quality Management Representative QA Division