



**Headoffice:** C-4th&5th, 16, Deogyong-daero 1556beon-gil,  
Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA  
**Manufacturing site:** 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu,  
Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499  
<http://www.sdbiosensor.com>

## Certificate of Analysis

<b>Product Name</b>		<b>STANDARD™ M10 Module</b>	
<b>Ref No.</b>	<b>M10-01M</b>	<b>Cat No.</b>	<b>11M1012</b>
<b>Lot No.</b>	<b>8A241J1AC1</b>	<b>MFG Date</b>	<b>2023.01.18.</b>
<b>Testing categories</b>		<b>Specification</b>	<b>Result</b>
<b>ANALYZER</b>	Apperance Inspection	No scratch, stain and crack etc.	<i>Pass</i>
		Top case assembly condition	<i>Pass</i>
	Function	Power Inspection	<i>Pass</i>
		Door open Function	<i>Pass</i>
		Fan Motor Function	<i>Pass</i>
		Ethernet Test	<i>Pass</i>
		Cartridge Recognition Test	<i>Pass</i>
F/W version check	<i>Pass</i>		
<b>PACKING</b>	Packaging Inspection	Analyzer / User guide / Accessory / Inner box	<i>Pass</i>
		Serial No. between analyzer label and package label	<i>Pass</i>
		Check LOT No. between package label	<i>Pass</i>

*According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.*

**SD Biosensor, Inc.**

Ko Jongkwan   
Quality Management Representative  
QA Division

Ko Jongkwan  
Quality Management Representative  
QA Division

Issue date: 2023-01-19