



**Headoffice:** C-4th&5th, 16, Deogyong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA  
**Manufacturing site:** 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA / 18-29, Cheomdangjeop 2-ro, Sandong-eup, Gumi-si, Gyeongsangbuk-do, 39174, Republic of Korea

TEL : +82-31-300-0400 / FAX : +82-31-300-0499  
<http://www.sdbiosensor.com>

## Certificate of Analysis

<b>Product Name</b>	<b>STANDARD™ M10 MDR-TB</b>		
<b>Ref No.</b>	<b>M10-MTB-01</b>	<b>Cat No.</b>	<b>11MTB10A</b>
<b>Lot No.</b>	8C44121AJ1	<b>MFG Date</b>	2023.01.02.
		<b>EXP Date</b>	2024.01.01.
<b>Testing categories</b>	<b>Specification</b>		<b>Result</b>
Performance Inspection	• Sensitivity: 100% detected in 2X LoD		Pass
	• Repeatability: Results are within ± 2 Ct of target concentration		Pass
	• Specificity: No fluorescence signal detected in negative control		Pass
Apperance Inspection	• Check the text number of the labeling material		Pass
	• Quantity per one container : 10ea		Pass
	• Expiration date between cartridge label and package of test kit • LOT No. between cartridge label and package of test kit		Pass

*According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.*

**SD Biosensor, Inc.**

Ko Jongkwan

Quality Management Representative  
 QA Division

\_\_\_\_\_  
 Jongkwan Ko  
 Quality Management Representative  
 QA Division

Issue date: 2023-01-02