

## **Declaration of Conformity**



Manufacturer Name SD Biosensor, Inc.

Manufacturer Address <u>Head Office</u>

C-4th&5th, 16, Deogyeong-daero, 1556beon-gil, Yeongtong-gu,

Suwon-si, Gyeonggi-do 16690, KOREA

Manufacturing Site

74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do 28161, KOREA

EC Representative Name MT Promedt Consulting GmbH

EC Representative Address Altenhofstrasse 80 66386 St. Ingbert Germany

Common Name Immunoassay Test Kit

Product Name STANDARD<sup>TM</sup> F COVID-19 Ag FIA

STANDARD<sup>TM</sup> COVID-19 Ag Control

\*Please refer to "Annex I. Product List" on page 2 in more detail.

**Reference Number** F-NCOV-01G, C-NCOV-01G

Catalog Number 10COV30D, 10COV31D, 10COVC10

Classification Others not covered by Annex II and self-testing according to

Directive 98/79/EC

Conformity Assessment Route Annex III of Directive 98/79/EC (EC Declaration of Conformity)

**Applied Standards** EN ISO 13485:2016 EN ISO 18113-1:2011

EN ISO 14971:2012 EN ISO 18113-2:2011 EN ISO 23640:2015 EN ISO 15223-1:2016 EN ISO 17511:2003 EN 62366:2008

EN 13612:2002

We herewith declare that the above mentioned products meet the provisions of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentations are retained under the premises of the manufacturer. SD Biosensor, Inc. is exclusively responsible for the declaration of conformity.

Place: Suwon-si, Republic of Korea Valid from: September 23, 2021

Signature

Hyo-Keun, Lee CEO / President



## Annex I. Product List

F-NCOV-01G STANDARD <sup>TM</sup> F COVID-19 Ag FIA (Catalog No. 10COV30D) - Test Device (individually in a foil pouch with desiccant) - Extraction buffer tube - Nozzle cap - Sterile swab – (Nasopharyngeal)	<b>EDMA Code</b> 15 04 90 90 00	<b>Description of EDMA code</b> Other Other Virology Reagents
F-NCOV-01G STANDARD <sup>TM</sup> F COVID-19 Ag FIA (Catalog No. 10COV31D) - Test Device (individually in a foil pouch with desiccant) - Extraction buffer tube - Nozzle cap - Sterile swab – (Nasal)	<b>EDMA Code</b> 15 04 90 90 00	<b>Description of EDMA code</b> Other Other Virology Reagents
C-NCOV-01G STANDARD <sup>TM</sup> COVID-19 Ag Control (Catalog No. 10COVC10) - STANDARD <sup>TM</sup> COVID-19 Ag Control - Positive - STANDARD <sup>TM</sup> COVID-19 Ag Control - Negative	EDMA Code 15 50 01 04 00	<b>Description of EDMA code</b> Other Virology Controls - Inf.Imm.