Headoffice: C-4th&5th, 16, Deogyeong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499
http://www sdbiosensor.com

Certificate of Analysis

Product Name	STANDARD F iFOB Control		
Cat No.	10IFOC10	Ref No.	C-F-FOB
Main Lot	9114C51AC	MFG Date	2023.12.05.
		EXP Date	2025.06.04.

Testing categories	Specification	Result
Appearance Inspection	• No tablet damage (Tablet crack, powder flying etc.) • No bottle scratch, stain and crack etc. • Control bottle and cap must be combined without gaps	*Pass*
Barcode Check	The product name and 'apply sample' should be displayed on the screen.	*Pass*
Repeatability	CV ≤ 10%	*Pass*
Control range	• Result within Level 1 control range • Result within Level 2 control range	*Pass*
Packaging Inspection	1 Level 1 Control bottle(10ea), 1 Level 2 Control bottle(10ea), 1 IFU	*Pass*
	Expiration date between the Control tablet and package of Control kit	*Pass*

For overseas use only

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.
Ko Jongkwan 고종관
Quality Management Representative
QA Division

Jongkwan Ko
Quality Management Representative
QA Division