

Importer Address: Plot No.38, Sector-4, IMT Manesar, Gurgaon, Haryana 122050, India

TEL: +82-31-300-0400 / FAX: +82-31-300-0499 http://www.sdbiosensor.com

## **Certificate of Analysis**

Product Name	STANDARD™ F U-Albumin FIA			
Cat No.	10UAL10B	Ref No.	F-UALB	
LOT No.	63947K1AC	MFG Date	2023.07.19.	
		EXP Date	2025.01.18.	
Testing categories	Specification		Result	
Appearance Inspection	<ul> <li>Uniform Gap between upper and lower device</li> <li>No scratch, stain and crack etc.</li> <li>No contamination</li> <li>There should be colored lines.</li> <li>The tablet must not be damaged.</li> <li>Check device type</li> <li>Check buffer amount</li> </ul>			Pass
Barcode Check	The product name and 'apply sample' should be displayed on the screen.			Pass
Repeatability	• Urine CV ≤ 10% EISEAS USE ON V			Pass
Dynamic range	<ul> <li>Testing temperature range: 15-32°C   59-90°F</li> <li>Testing Range: 5.0 ~ 250 mg/L</li> </ul>			Pass
Packaging Inspection	20 Test Device, 20 Buffer, 1 IFU			Pass
	No leackage in sealing part of pouch			Pass
	No leackage in sealing part of buffer solution tube		Pass	
	Manufacturing and expiration date between the pouch and package of test kit.			Pass
	LOT No. between the pouch and package of test kit.			Pass

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc.

Ko Jongkwan

Quality Management Representative **QA** Division

Ko Jongkwan

Quality Management Representative

QA Division

Issue date: 2023.08.25.